



# CERTIFICATE OF ACCREDITATION

**ANSI-ASQ National Accreditation Board**

500 Montgomery Street, Suite 625, Alexandria, VA 22314, 877-344-3044

This is to certify that

**Teledyne LeCroy  
700 Chestnut Ridge Road  
Chestnut Ridge, NY 10977**

has been assessed by ANAB  
and meets the requirements of international standard

**ISO/IEC 17025:2005**

and national standard

**ANSI/NCSL Z540-1-1994 (R2002)**

while demonstrating technical competence in the field of

**CALIBRATION**

Refer to the accompanying Scope of Accreditation for information regarding the types of calibrations to which this accreditation applies.

AC-2555

Certificate Number

  
ANAB Approval

Certificate Valid: 02/14/2018-02/14/2020  
Version No. 001      Issued: 02/14/2018



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



ANSI-ASQ National Accreditation Board

## SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005 AND ANSI/NCSL Z540-1-1994 (R2002)

### Teledyne LeCroy

700 Chestnut Ridge Road

Chestnut Ridge, NY 10977

Paul Mowat 845-578-6060

paul.mowat@teledyne.com www.teledyne.com

### CALIBRATION

Valid to: February 14, 2020

Certificate Number: AC-2555

#### Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Source	(0 to 330) mV (0.33 to 3.3) V (3.3 to 33) V (33 to 330) V (330 to 1 000) V	47 $\mu\text{V}/\text{V} + 3.3 \mu\text{V}$ 40 $\mu\text{V}/\text{V} + 4.5 \mu\text{V}$ 40 $\mu\text{V}/\text{V} + 39 \mu\text{V}$ 43 $\mu\text{V}/\text{V} + 0.39 \text{mV}$ 43 $\mu\text{V}/\text{V} + 1.2 \text{mV}$	Fluke 5500A Multi Product Calibrator
DC Voltage – Measure <sup>1</sup>	(10 to 100) mV 100 mV to 1 V (1 to 10) V (10 to 100) V (100 to 1 000) V	9 $\mu\text{V}/\text{V} + 1.2 \mu\text{V}$ 5.5 $\mu\text{V}/\text{V} + 1.2 \mu\text{V}$ 5.4 $\mu\text{V}/\text{V} + 1.9 \mu\text{V}$ 8.1 $\mu\text{V}/\text{V} + .041 \text{mV}$ 9.7 $\mu\text{V}/\text{V} + 0.31 \text{mV}$	HP 3485A w/ option 002 Multimeter
DC Current – Source	(0 to 3.2) mA (3.3 to 32) mA (33 to 320) mA 330mA to 2.1 A (2.2 to 11) A	0.10 mA/A + 0.04 $\mu\text{A}$ 80 $\mu\text{A}/\text{A} + 0.22 \mu\text{A}$ 82 $\mu\text{A}/\text{A} + 2.8 \mu\text{A}$ 0.24 mA/A + 34 $\mu\text{A}$ 0.47 mA/A + 0.26 mA	Fluke 5500A Multi Product Calibrator
DC Current – Measure	(10 to 100) $\mu\text{A}$ (0.1 to 1) mA (1 to 10) mA (10 to 100) mA (0.1 to 1) A	28 $\mu\text{A}/\text{A} + 0.81 \text{nA}$ 26 $\mu\text{A}/\text{A} + 5.6 \text{nA}$ 26 $\mu\text{A}/\text{A} + 60 \text{nA}$ 42 $\mu\text{A}/\text{A} + 0.52 \mu\text{A}$ 0.12 mA/A + 10 $\mu\text{A}$	HP 3485A w/ option 002 Multimeter



**ANSI-ASQ National Accreditation Board**

**ANAB**

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Source	(0 to 11) Ω (11 to 33) Ω (33 to 330) Ω (0.33 to 3.3) kΩ (3.3 to 33) kΩ (33 to 110) kΩ (110 to 330) kΩ (0.33 to 3.3) MΩ (3.3 to 11) MΩ (11 to 33) MΩ (33 to 110) MΩ (110 to 330) MΩ	0.012 % + 6.2 mΩ 0.010 % + 0.012 Ω 0.007 % + 0.012 Ω 0.007 % + 0.05 Ω 0.007 % + 0.47 Ω 0.009 % + 4.7 Ω 0.01 % + 4.7 Ω 0.012 % + 43 Ω 0.047 % + 0.43 kΩ 0.1 % + 0.43 kΩ 0.4 % + 4.3 kΩ 0.4 % + 13 kΩ	Fluke 5500A Multi Product Calibrator  (2-wire mode from 110 kΩ to 330 MΩ)
Resistance – Measure <sup>1</sup>	0 to 10 Ω (10 to 100) Ω (0.1 to 1) kΩ (1 to 10) kΩ (10 to 100) kΩ (0.1 to 1) MΩ (1 to 10) MΩ (10 to 100) MΩ	18 μΩ/Ω + 51 μΩ 18 μΩ/Ω + 0.5 mΩ 13 μΩ/Ω + 0.5 mΩ 13 μΩ/Ω + 5.2 mΩ 13 μΩ/Ω + 0.05Ω 17 μΩ/Ω + 2.0 Ω 53 μΩ/Ω + 100 Ω 0.055 % + 1 kΩ	HP 3458A w/ option 002 Multimeter
AC Voltage – Source	(1 to 33) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (33 to 330) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.3 % + 20 μV 0.12 % + 16 μV 0.16 % + 16 μV 0.19 % + 16 μV 0.27 % + 26 μV 0.79 % + 47 μV  0.22 % + 39 μV 0.04 % + 16 μV 0.08 % + 16 μV 0.12 % + 38 μV 0.19 % + 0.13mV 0.54 % + 0.26mV	Fluke 5500A Multi Product Calibrator



**ANSI-ASQ National Accreditation Board**

**ANAB**

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source	(0.33 to 3.3) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz  (3.3 to 33) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz  (33 to 330) V 45 Hz to 1 kHz (1 to 10) kHz (10 to 20) kHz  (330 to 1 020) V 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.15 % + 0.2mV 0.02 % + 47 µV 0.06 % + 47 µV 0.06 % + 47 µV 0.19 % + 1.3 mV 0.39 % + 2.6 mV  0.12 % + 2.0 mV 0.03 % + 0.47 mV 0.062 % + 2.0 mV 0.15 % + 4.0 mV 0.19 % + 13 mV  0.04 % + 5 mV 0.06 % + 12 mV 0.07 % + 26 mV  0.04 % + 62mV 0.16 % + 77 mV 0.16 % + 0.39V	Fluke 5500A Multi Product Calibrator
AC Voltage – Measure <sup>1</sup>	(10 to 100) mV 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz  100 mV to 1 V 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz  (1 to 10) V 40 Hz to 1 kHz (1 to 10) kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.34mV/V +1.1µV 0.41mV/V +1.1µV 1.3mV/V + 1.1µV 5.0mV/V + 1.1µV  90µV/V + 20µV 0.16mV/V + 20µV 0.32mV/V + 20µV 0.82mV/V +29µV  85µV/V + 0.2mV 0.15mV/V +0.2mV 0.15mV/V +0.2mV 0.33mV/V + 0.6mV 0.81mV/V +2.3mV	HP 3458A Multimeter



**ANSI-ASQ National Accreditation Board**

**ANAB**

### Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(10 to 100) V 40 Hz to 1 kHz (1 to 10) kHz (10 to 20) kHz (100 to 700) V 40 Hz to 1 kHz (1 to 5) kHz (10 to 20) kHz	0.22mV/V + 2mV 0.23mV/V + 2mV 0.23V/V + 2.0mV  0.41mV/V + 20mV 0.60mV/V + 20mV 0.61mV/V + 66mV	HP 3458A Multimeter
Oscilloscopes <sup>1</sup> – Frequency Ref - Lo BW	10 MHz	0.25 $\mu$ Hz/Hz	HP 8648C w/option 1E5 Signal Generator
Frequency Ref – Hi BW	10 MHz	50 nHz/Hz	Anritsu Generator phase locked to Stanford Research Systems Rb Frequency Standard PSR10
Bandwidth	(-20 to +20) dBm 10 kHz to 4 GHz  (-20 to +20) dBm 25 MHz to 50 GHz 25 MHz to 65 GHz	0.4 dB  0.64 dB 0.90 dB	Agilent 4418B Power Meter w/E9304 H18 Power Sensor  R&S NRPZ57 Power Sensor

### Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure	(-60 to +20) dBm 9 kHz to 4 GHz	0.40 dB	Agilent 4418B Power Meter w/E9304 H18 Power Sensor
	(-35 to +20) dBm DC to 50 GHz (50 to 65) GHz	0.45 dB 0.61 dB	R&S NRPZ57 Power Sensor



ANSI-ASQ National Accreditation Board

**ANAB**

### Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Measure	1 kHz to 300 MHz	0.5 nHz/Hz + 0.58 $\mu$ Hz	Stanford Research Systems - SR620 Counter, phase locked to Rb Frequency Standard PSR10
Frequency – Generate	10 MHz	50 nHz/Hz	Stanford Research Systems Rb Frequency Standard - PSR10

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2555.



Vice President

